

**CO<sub>2</sub> POD**

**CONNECTIONS**

Sensor Type	Serial coding	Connections			
		A	B	C	D
CO <sub>2</sub>	Cxxxx		green	red	blue

Cable colours as shown relate to black outer cable.  
With white outer cable, connect yellow cable to B, others as above.



Item  
10G50

**CALIBRATION**

None

**WEB**

On [www.general-alert.com](http://www.general-alert.com), login using Username and Password.

**Options:**

- Name the POD (CO<sub>2</sub> Lab 3 etc.)
- Set Alarm levels (above or below specific CO<sub>2</sub> levels)
- Establish times of day and days of week for Alarm settings
- Enable Daily Data Feed (for online graphing and recording)
- Establish Users to be notified on POD Alerts
- Obtain instant live Status Report from POD

**MOBILE**

To access features you need to be a listed User on a General Alert Account.

**Features:**

Receive Alerts

'CO<sub>2</sub> (C5001) Alert: CO<sub>2</sub> = 2555 ppm on Alarm 2 (Normal = 2000ppm)

Cancel Alarm (while the problem is being sorted)

5001.0.2

Request Status Report for live data

5001.6

Receive Status Report

'CO<sub>2</sub> Lab 3 (C0005001).  
Reading = 2698 ppm at Dec 25 2012 16.45 (local time)

PO Box 103, Safron Walden, CB11 3GD, UK  
T: +44 1799 550837, e: [info@general-alert.com](mailto:info@general-alert.com)  
W: [www.general-alert.com](http://www.general-alert.com)