


pH POD

Item
10G10

CONNECTIONS

Sensor Type	Serial coding	Connection
pH	Pxxxx	BNC



CALIBRATION

Calibration is done following on-screen instructions.

Options / Pod Edit / choose POD / Calibrate / use buffer 7 then 4.

Note – Stir the buffer solutions before Pressing OK for each of the Buffer 7 and Buffer 4.

WEB

On www.general-alert.com, login using Username and Password.

Options:

- Name the POD (Tank A, Fermentation 23 etc.)
- Set Alarm levels (above or below specific pH levels)
- Establish times of day and days of week for Alarm settings
- Enable Daily Data Feed (for online graphing and recording)
- Establish Users to be notified on POD Alerts
- Obtain instant live Status Report from POD

MOBILE

To access features you need to be a listed User on a General Alert Account.

Features:

Receive Alerts

'pH Tank A (T1098) Alert:
pH = 4.98 on Alarm 2
(Normal = 4.97)

Cancel Alarm (while the problem
is being sorted)

1098.0.2

Request Status Report for live data

1098.6

Receive Status Report

'pH Tank A (T1098).
Reading = 4.95 at Aug 08
2010 16.45 (local time)

PO Box 103, Safron Walden, CB11 3GD, UK
T: +44 1799 550837, e: info@general-alert.com
W: www.general-alert.com